

AD-A190 659



TECHNICAL REPORT BRL-TR-2856

A 150,000 POUNDS PER SQUARE INCH
DYNAMIC PRESSURE CALIBRATOR

ARPAD A. JUHASZ, DONALD H. NEWHALL
CHARLES D. BULLOCK, JAMES O. PILCHER
MELINDA B. KRUMMERICH

OCTOBER 1987

APPROVED FOR PUBLIC RELEASE; DISTRIBUTION UNLIMITED.

US ARMY BALLISTIC RESEARCH LABORATORY
ABERDEEN PROVING GROUND, MARYLAND

DESTRUCTION NOTICE

Destroy this report when it is no longer needed. DO NOT return it to the originator.

Additional copies of this report may be obtained from the National Technical Information Service, U.S. Department of Commerce, Springfield, VA 22161.


The findings of this report are not to be construed as an official Department of the Army position, unless so designated by other authorized documents.

The use of trade names or manufacturers' names in this report does not constitute indorsement of any commercial product.

UNCLASSIFIED

SECURITY CLASSIFICATION OF THIS PAGE

REPORT DOCUMENTATION PAGE				Form Approved OMB No. 0704-0188 Exp. Date: Jun 30, 1986	
1a. REPORT SECURITY CLASSIFICATION Unclassified			1b. RESTRICTIVE MARKINGS		
2a. SECURITY CLASSIFICATION AUTHORITY			3. DISTRIBUTION/AVAILABILITY OF REPORT		
2b. DECLASSIFICATION/DOWNGRADING SCHEDULE					
4. PERFORMING ORGANIZATION REPORT NUMBER(S) BRL-TR-2856			5. MONITORING ORGANIZATION REPORT NUMBER(S)		
6a. NAME OF PERFORMING ORGANIZATION US Army Ballistic Rsch Lab		6b. OFFICE SYMBOL (If applicable) SLCBR-IB	7a. NAME OF MONITORING ORGANIZATION		
6c. ADDRESS (City, State, and ZIP Code) Aberdeen Proving Ground, MD 21005-5066			7b. ADDRESS (City, State, and ZIP Code)		
8a. NAME OF FUNDING /SPONSORING ORGANIZATION		8b. OFFICE SYMBOL (If applicable)	9. PROCUREMENT INSTRUMENT IDENTIFICATION NUMBER		
8c. ADDRESS (City, State, and ZIP Code)			10. SOURCE OF FUNDING NUMBERS		
			PROGRAM ELEMENT NO.	PROJECT NO.	TASK NO.
					WORK UNIT ACCESSION NO.
11. TITLE (Include Security Classification) A 150,000 POUNDS PER SQUARE INCH DYNAMIC PRESSURE CALIBRATOR					
12. PERSONAL AUTHOR(S) Juhasz, Arpad A., Newhall, Donald H., Bullock, Charles D., Pilcher, James O., Krummerich, M.B.					
13a. TYPE OF REPORT TR		13b. TIME COVERED FROM _____ TO _____		14. DATE OF REPORT (Year, Month, Day)	
				15. PAGE COUNT	
16. SUPPLEMENTARY NOTATION					
17. COSATI CODES			18. SUBJECT TERMS (Continue on reverse if necessary and identify by block number)		
FIELD	GROUP	SUB-GROUP			
19. ABSTRACT (Continue on reverse if necessary and identify by block number) A positive step pressure generator suitable for the dynamic calibration of ballistic pressure transducers is described. The device uses a hydraulic system to generate pressures up to 150,000 pounds per square inch and is capable of delivering pressure steps in one millisecond or less. Final pressure step values can be traced to conventional deadweight calibration figures with an accuracy of 0.2% by using a transfer standard (static reference gage). Among the uses for the device are examining transducer response characteristics to rapid positive pressure steps and comparing the relative response behavior of several transducers to a common pressure step.					
20. DISTRIBUTION/AVAILABILITY OF ABSTRACT <input checked="" type="checkbox"/> UNCLASSIFIED/UNLIMITED <input type="checkbox"/> SAME AS RPT. <input type="checkbox"/> DTIC USERS			21. ABSTRACT SECURITY CLASSIFICATION Unclassified		
22a. NAME OF RESPONSIBLE INDIVIDUAL Arpad A. Juhasz			22b. TELEPHONE (Include Area Code) (301) 278-6158		22c. OFFICE SYMBOL SLCBR-IB-B



ACKNOWLEDGEMENTS

The authors gratefully acknowledge the contributions of the staff at Harwood Engineering Corporation for engineering design and construction of the step calibrator. Thanks are also extended to Mr. W. J. Donovan, Ballistic Research Laboratory, APG, MD and Mr. J. D. Dykstra, CSTA, APG, MD, for technical discussions relating to the project.

TABLE OF CONTENTS

	<u>Page</u>
LIST OF ILLUSTRATIONS.....	7
I. INTRODUCTION	
A. Ballistic Pulse Method.....	9
B. Shock Tube Method.....	9
C. Negative-Going Pressure Step Method.....	9
D. Positive-Going Pressure Step Method.....	10
II. DISCUSSION	
A. Design Concept.....	10
B. Device Description.....	13
C. Device Operation.....	15
D. Prototype Performance.....	15
III. CONCLUSIONS.....	20
REFERENCES.....	21
DISTRIBUTION LIST.....	23

LIST OF ILLUSTRATIONS

Figure		Page
1	Design Concept for the Positive Step Pressure Calibrator.....	11
2	Engineering Schematic of 150,000 psi Positive Step Pressure Calibrator.....	12
3	Lateral Cross-Sectional View of Test Head.....	13
4	Closeup View of Test Chamber After Trigger Actuation.....	14
5	Pressure-Time History of a 75,000 psi Dynamic Pressure Step; 10 second Window.....	16
6	Pressure-Time History of a 75,000 psi Dynamic Pressure Step; 20 millisecond Window.....	16
7	Pressure-Time History of a 75,000 psi Dynamic Pressure Step; 5 millisecond Window.....	17
8	Pressure-Time History of a 75,000 psi Dynamic Pressure Step; 2 millisecond Window.....	17
9	Response of a Piezoelectric Ballistic Pressure Gage to a 100,000 psi Positive Pressure Step.....	18
10	Response of a Strain Type Ballistic Pressure Gage to a 100,000 psi Positive Pressure Step.....	18
11	Response of a Developmental Pressure Gage to a 125,000 psi Positive Pressure Step.....	19
12	Response of a Faulty Developmental Pressure Gage to a 130,000 psi Positive pressure Step.....	19

I. INTRODUCTION

Although ballistic pressure transducers (gages) are used to measure dynamic events which occur in milliseconds, the determination of their response characteristics is routinely limited to static calibration against a deadweight pressure standard. The strength of this method is that the deadweight device is a primary standard. Its weakness lies in the assumption that the static and dynamic responses of the gage in question are equivalent. Differences in gage response to static and dynamic events, however, can lead to serious measurement errors. There has been general agreement in the measurement community that dynamic techniques are needed to supplement current static calibration methods. Several techniques have been developed to address this problem.

1. BALLISTIC PULSE METHOD

In one version of this technique, the gage is mounted at the end of a tube in contact with a hydraulic fluid confined by a movable piston. The tube guides a projectile which impacts the piston to create a pressure pulse in the fluid. Different pressures may be achieved by varying the compressibility of the fluid, the mass of the piston, and the mass and velocity of the projectile. The pulses rise within milliseconds and mimic the characteristic rising and falling of a ballistic pressure pulse. One such device, capable of operating to a pressure of 100,000 pounds per square inch, is operational at the Combat Systems Testing Activity (CSTA), Aberdeen Proving Ground, Maryland.³

This method is quite useful for dynamic comparison of several different pressure gages; however, variations in projectile velocity, frictional effects on the moving piston, and other energy losses make it difficult to accurately compute the actual delivered pressures. Because a projectile is fired during the calibration process, this method requires more extensive safety provisions than are readily available in most laboratories.

2. SHOCK TUBE METHOD

Two general approaches of shock tube calibration are followed. In the first, the test gage is mounted in the end wall of a tube and subjected to a reflecting shock wave. The gage output is monitored as the shock front arrives at and reflects from the end wall. In the second approach, the gage is mounted in the side wall of the tube and its output is monitored as the shock front passes. Both methods generate rapidly rising pressure pulses that are readily calculated by thermodynamic principles using velocity measurements and gas properties.

Shock tube methods are useful in establishing the dynamic response characteristics of pressure gages. However, calibration is generally limited to pressures below 1000 pounds per square inch, whereas ballistic applications require far higher pressures.

3. NEGATIVE-GOING PRESSURE STEP METHOD

In this technique, the gage is exposed to a given pressure under static conditions using a hydraulic fluid. The pressure on the gage is then

relieved using a fast acting dump valve, bringing the system to atmospheric pressure. The gage output obtained during the depressurization is assumed to be the inverse of the corresponding positive pressure step.

This method's strengths include its relative simplicity and suitability for use in calibration facilities. The response of the negative step calibrator can be very quick, 100 microseconds or less. However, the major assumption, that the positive response of the gage is equal and opposite to the negative response of the gage, is not completely accurate; pressure preloading of the gage-to-mount interface and hysteresis can cause significant differences between the pressurization and depressurization pulses.

4. POSITIVE-GOING PRESSURE STEP METHOD

In this technique, the gage, initially at atmospheric pressure, is subjected to a pressure increase by the opening of a fast-acting valve. Because the final pressure value is held, the method is suitable for obtaining calibration response data. Although this technique holds a great deal of promise, the engineering details of creating a working device can be formidable. Johnson and Cross of the National Bureau of Standards had designed a 50,000 pounds per square inch step calibrator in the 1950's. Smith and Dykstra described low pressure versions of this device. These calibrators were successfully used to generate positive pressure steps up to 5000 pounds per square inch in less than one millisecond.

Building on these ideas, we have developed a device capable of generating precisely known positive pressure steps up to 150,000 pounds per square inch in less than one millisecond. The step calibrator may be safely operated in a laboratory environment and can compare the response of several transducers to a common pressure step. Additionally, it is economical to use as a routine laboratory tool for gage calibration and screening.

The discussion that follows describes the pressure step calibrator and presents several examples of its operation.

II. DISCUSSION

1. DESIGN CONCEPT

Figure 1 illustrates the basic configuration of the device, details are given in Figure 2. A large pressure reservoir is connected to a much smaller test chamber by a fast acting ball valve. The test chamber is equipped with several gage ports and a vacuum port which aids in filling the reservoir and setting the baseline chamber pressure. The large ratio of reservoir to test chamber free volume reduces overall system pressure drop while generating the pressure step. Reservoir pressure is provided by a conventional hydraulic high pressure panel and monitored by a high quality static reference gage. The maximum operating design pressure is 150,000 pounds per square inch and the specified action time (10% - 90% of peak pressure) is under one millisecond.

Outputs of both the static reference gage and the ballistic test gages are monitored during the course of the test. The final steady state output of the reference gage is taken as the true value of the pressure step maximum. The short- and long-term monitoring of the test gage outputs establish the relationship between dynamic and steady state response behavior.




Figure 1. Design Concept for the Positive Step Pressure Calibrator




Figure 2. Engineering Schematic of 150,000 psi Positive Step Pressure Calibrator

- | | |
|------------------------|---------------------------|
| (1) Pressure reservoir | (10) Piston guide bushing |
| (2) Connecting channel | (11) Pin joints |
| (3) Test chamber | (12) Trigger mechanism |
| (4) Ball valve | (13) Hydraulic jack |
| (5) Valve seat | (14) Limit/stop buffer |
| (6) Gage port | (15) Vacuum stem valve |
| (7) Vacuum line port | (16) Port |
| (8) Test head | (17) High pressure source |
| (9) Valve actuator | |

2. DEVICE DESCRIPTION

The device, illustrated in Figure 2 thru 4, consists of a large pressure reservoir (1) opening into a short wide channel (2) which terminates in a very small cylindrical test chamber (3). Located in the test chamber (3) is a ball valve (4) which provides a high pressure seal at either of the valve seats (5) located at each end of the test chamber (3). Located in the side wall of the test chamber (3) are four gage ports (6) and one vacuum line port (7). The channel (2), test chamber (3), gage ports (6) and vacuum line port (7) are contained in the test head (8), a monolithic assembly shown in lateral cross section in Figure 3. The test head (8) is readily removed from the assembly, allowing changing of the gage ports (6) and the test chamber (3). The ratio of the reservoir (1) volume to the test chamber (3) free volume is 197:1. The channel (2) is kept short and wide to minimize retardation of fluid flow during the operation cycle. The end closure of the test chamber (3) is formed by the ball valve actuator piston (9) and the piston guide bushing (10). The piston is actuated by a quick release top-dead-center mechanism which consists of three pin joints (11), an air controlled trigger mechanism (12), a hydraulic jack (13) and a limit stop/buffer (14). The system shown in Figure 2 is in the cocked position with the ball (4) pressed against the seat (5) that isolates the test chamber (3) from the reservoir (1). The jack (13) is pressurized to provide sufficient force to seal the reservoir (1) from the test chamber (3). The ratio of reservoir (1) pressure to jack (13) pressure is approximately 100:1.

A hollow stem valve (15) connected to a vacuum/drain line is located at port (7). See Figure 3.




Figure 3. Lateral Cross-Sectional View of Test Head

A high pressure line (17), (Figure 2), is connected to port (16) at the upper end of the reservoir (1). This line connects the pressure generation system to the measurement system.

Prior to operation, both the test chamber and reservoir are filled with fluid at ambient pressure. During pressurization the reservoir is sealed off from the test chamber. The actual movement of fluid through channel (2) is minimal, consisting of the volume required to displace the movement of actuator piston (9) in the test chamber (.002 cc) and the amount due to compression of the fluid originally in the test chamber.

Figure 4 shows the device after the trigger mechanism (12) is released. The trigger (12) forces the middle pin joint of the top-dead-center mechanism against the limit stop/buffer (14), relaxes the force generated by the jack (13) and withdraws the ball valve actuator piston (9) into the piston guide bushing (10). Differential pressure between the reservoir (1) and the test chamber (3) forces the valve ball (4) against the lower seat (5) on the piston guide bushing (9). This action forms a new seal at the piston end of the test chamber (3) and allows fluid to flow from the reservoir (1) to the test chamber (3) causing the pressure in the chamber to rise to approximately 98% of the original reservoir pressure.




Figure 4. Closeup View of Test Chamber After Trigger Actuation

3. DEVICE OPERATION

Starting with a drained system, the top-dead-center mechanism is placed in the release position as shown in Figure 4. The high pressure line (17) is closed off and the vacuum valve (15) is opened. Gages are mounted in the gage ports (6), the system is evacuated to a pressure of 2 Torr and the vacuum valve (15) is closed. A 50% solution of water and glycol with a rust inhibitor enters the system through the high pressure line (17). Liquid is used rather than gas to minimize the level of stored energy. When the system is filled and stabilized at atmospheric pressure, the top-dead-center mechanism is cocked as shown in Figure 2. The hydraulic jack (13) is pressurized to approximately 1% of the desired reservoir pressure. Monitored by the static reference gage, the pressure generation system pressurizes the reservoir (1). Once the desired reservoir pressure has been established, the high pressure line (17) is closed off from the standard gage and the pressure generation system by a constant volume valve. The device is now ready to be triggered.

Activation of the top-dead-center mechanism shown in Figure 4 initiates the event and the recording system. Pre-trigger delay features permit the recording of initial baseline pressures, the rising portion and final steady state values of the pressure-time curve. One gage with exceptionally good response characteristics and known history is used as an informal laboratory standard. The output from this gage is monitored for at least 10 seconds after the trigger event to observe system behavior, particularly possible pressure losses from leakage. Soon after the system is triggered, the constant volume valve to the static reference gage is reopened and the reservoir pressure measured. Thus, the speed of the step can be measured using the timebase of the recording system and the magnitude of the dynamic response can be checked against the response of the static reference gage (a secondary standard).

At the completion of the test, the top-dead-center mechanism is recocked, the test chamber (3) is drained through the vacuum valve (15), and the pressure is relaxed in the reservoir (1). The gages can now be replaced for further testing.

4. PROTOTYPE PERFORMANCE

Examples of both short- and long-term responses of a commercial piezoelectric pressure gage to a positive-going pressure step are presented in Figures 5 thru 8. Figure 5 shows a typical pressure versus time history for a 75,000 pounds per square inch pulse acquired over 10 seconds. Monitoring the test gage response for 10 seconds allows correlation to the steady state response of the static reference gage. Figure 6 shows a 20 millisecond window of the same 75,000 pounds per square inch pulse; Figures 7 and 8 present progressively shorter windows. Overlaying the traces indicates that the steady state value of the pressure step is rapidly achieved and held after initial oscillations have died out. Similar oscillations have been reported in lower pressure devices by both Smith and Dykstra; one may conclude that these oscillations are caused by actions within the pressure generation system and the mount, not by the gage itself.




Figure 5. Pressure-Time History of a 75,000 psi Dynamic Pressure Step; 10 second Window




Figure 6. Pressure-Time History of a 75,000 psi Dynamic Pressure Step; 20 millisecond Window




Figure 7. Pressure-Time History of a 75,000 psi Dynamic Pressure Step; 5 millisecond Window




Figure 8. Pressure-Time History of a 75,000 psi Dynamic Pressure Step; 2 millisecond Window

One important application of the step calibrator is comparing the responses of different types of pressure gages to the same input. Figures 9 and 10 show the pressure versus time histories of both a piezoelectric gage and a strain-type gage measuring a positive-going 100,000 pounds per square inch pressure step. The traces are virtually identical, exhibiting initial system oscillations which quickly decay.




Figure 9. Response of a Piezoelectric Ballistic Pressure Gage to a 100,000 psi Positive Pressure Step




Figure 10. Response of a Strain Type Ballistic Pressure Gage to a 100,000 psi Positive Pressure Step

The calibrator may also be used to analyze the behavior of experimental gages by comparing their output with a known standard. Figure 11 shows the pressure versus time history of a developmental pressure gage subjected to a 125,000 pounds per square inch pressure pulse. This curve exhibits, as expected, a smooth pressure rise and stable output after the peak pressure has been attained.




Figure 11. Response of a Developmental Pressure Gage to a 125,000 psi Positive Pressure Step




Figure 12. Response of a Faulty Developmental Pressure Gage to a 130,000 psi Positive Pressure Step

The positive step calibration device in conjunction with conventional deadweight calibration methods can be useful in tracing dynamic gage response problems. Figure 12 exhibits the pressure history of a developmental pressure gage exposed to a 130,000 psi pressure step. The trace indicates a clear upward drift after the step is complete. The same gage had exhibited good response behavior on static calibration using the deadweight system. In gun tests, however, the device read five percent low. From the dynamic calibration tests it became clear that it took tens of milliseconds for the gage output to reach a maximum. This could not be detected by the static calibration technique. In point of fact, on dissecting the gage it was found that the bond between the strain patch and the gage body was faulty. The defect of the type noted could well account for the differences in static vs dynamic behavior noted in testing.

III. CONCLUSIONS

The positive-going step calibration device described allows the accurate, safe and simple dynamic calibration and evaluation of ballistic pressure gages. The gage response obtained can be related to its static deadweight behavior. Several gages may be evaluated simultaneously relative to a common dynamic event. The calibrator may also be used as a diagnostic tool in analyzing and developing experimental pressure gages.

REFERENCES

1. National Bureau of Standards Report 4440, Measurement of High Pressure: Bibliography, Index and Preliminary Survey, W.G. Brombacher, U.S. Department of Commerce, National Bureau of Standards, October, 1955.
2. NBS Technical Note 914, A New Dynamic Pressure Source for the Calibration of Pressure Transducers, Carol F. Vezzetti, John S. Hilton, J. Franklin Mayo-Wells and Paul S. Lederer, U.S. Department of Commerce, National Bureau of Standards, June 1976. (and references therein)
3. Personal communication with J.D. Dykstra of the Combat Systems Testing Activity, Aberdeen Proving Ground, MD.
4. Smith, R.O., A Liquid Medium Step Function Pressure Calibrator, ASME Paper Number 63-WA-263. The American Society of Mechanical Engineers, 345 East 47th St. New York, NY.
5. Dykstra, J.D., "Evaluation of Pressure Transducer Response with a Pressure Step Generator," Bulletin of the Eleventh Meeting of the Joint Army-Navy-Air Force Solid Propellant Rocket Static Test Panel, SPIA Publication SPSTP/11, September, 1962.

DISTRIBUTION LIST

<u>No. Of</u> <u>Copies</u>	<u>Organization</u>	<u>No. Of</u> <u>Copies</u>	<u>Organization</u>
12	Commander Defense Technical Info Center ATTN: DTIC-DDA Cameron Station Alexandria, VA 22304-6145	1	Commander US Army Materiel Command ATTN: AMCDRA-ST 5001 Eisenhower Avenue Alexandria, VA 22333-0001
1	Commander USA Concepts Analysis Agency ATTN: D. Hardison 8120 Woodmont Avenue Bethesda, MD 20014	22	Commander Armament R&D Center US Army AMCCOM ATTN: SMCAR-TSS SMCAR-TDC SMCAR-LC LTC N. Barron SMCAR-AEE-B A. Beardell D. Downs S. Einstein S. Westley S. Bernstein P. Kemmey A. Bracuti J. Rutkowski SMCAR-LCB-I D. Spring SMCAR-LCE, R. Walker SMCAR-LCM-E S. Kaplowitz SMCAR-LCS SMCAR-LCU-CT E. Barrieres R. Davitt SMCAR-LCU-CV C. Mandala SMCAR-LCW-A M. Salsbury SMCAR-SCA L. Stiefel B. Brodman Picatinny Arsenal, NJ 07806-5000
1	HQDA/DAMA-ZA Washington, DC 20310		
1	HQDA, DAMA-CSM, E. Lippi Washington, DC 20310		
1	HQDA/DAMA-ART-M Washington, DC 20310		
1	HQDA/SARDA Washington, DC 20310		
1	Commander US Army War College ATTN: Library-FF229 Carlisle Barracks, PA 17013		
1	Director US Army BMD Advanced Technology Center P. O. Box 1500 Huntsville, AL 35807		
1	Chairman DOD Explosives Safety Board Room 856-C Hoffman Bldg. 1 2461 Eisenhower Avenue Alexandria, VA 22331		
1	Commander US Army Materiel Command ATTN: AMCPM-GCM-WF 5001 Eisenhower Avenue Alexandria, VA 22333	3	Project Manager Tank Main Armament System ATTN: AMCPM-TMA, K. Russell AMCPM-TMA-105 AMCPM-TMA-120 Picatinny Arsenal, NJ 07806-5000

DISTRIBUTION LIST

<u>No. Of Copies</u>	<u>Organization</u>	<u>No. Of Copies</u>	<u>Organization</u>
1	Commander US Army Watervliet Arsenal ATTN: SARWV-RD, R. Thierry Watervliet, NY 12189	1	Commander ERADCOM Technical Library ATTN: STET-L Fort Monmouth, NJ 07703-5301
4	Commander US Army Armament Munitions and Chemical Command ATTN: SMCAR-ESP-L L.Ambrosini AMSMC-IRC, G. Cowan AMSMC-LEM W. Fortune R. Zastrow Rock Island, IL 61299-7300	1	Commander US Army Harry Diamond Lab. ATTN: DELHD-TA-L 2800 Powder Mill Road Adelphi, MD 20783
1	Director Benet Weapons Laboratory Armament R&D Center US Army AMCCOM ATTN: SMCAR-LCB-TL Watervliet, NY 12189	1	Commander US Army Missile Command Rsch, Dev, & Engr Ctr ATTN: AMSMI-RD Redstone Arsenal, AL 35898
1	Commander US Army Aviation Research and Development Command ATTN: AMSAV-E 4300 Goodfellow Blvd. St. Louis, MO 63120	1	Director US Army Missile & Space Intelligence Center ATTN: AIAMS-YDL Redstone Arsenal, AL 35898-5500
1	Commander US Army TSARCOM 4300 Goodfellow Blvd. St. Louis, MO 63120	1	Commandant US Army Aviation School ATTN: Aviation Agency Fort Rucker, AL 36360
1	Director US Army Air Mobility Research And Development Laboratory Ames Research Center Moffett Field, CA 94035	1	Commander US Army Tank Automotive Command ATTN: AMSTA-TSL Warren, MI 48397-5000
1	Commander US Army Communications Electronics Command ATTN: AMSEL-ED Fort Monmouth, NJ 07703	1	Commander US Army Tank Automotive Command ATTN: AMSTA-CG Warren, MI 48090
		1	Project Manager Improved TOW Vehicle ATTN: AMCPM-ITV US Army Tank Automotive Command Warren, MI 48090

DISTRIBUTION LIST

<u>No. Of</u> <u>Copies</u>	<u>Organization</u>	<u>No. Of</u> <u>Copies</u>	<u>Organization</u>
1	Program Manager M1 Abrams Tank System ATTN: AMCPM-GMC-SA, T. Dean Warren, MI 48090	1	Commandant US Army Infantry School ATTN: ATSH-CD-CSO-OR Fort Benning, GA 31905
1	Project Manager Fighting Vehicle Systems ATTN: AMCPM-FVS Warren, MI 48090	1	President US Army Artillery Board Ft. Sill, OK 73503
1	President US Army Armor & Engineer Board ATTN: ATZK-AD-S Fort Knox, KY 40121	1	Commandant US Army Command and General Staff College Ft Leavenworth, KS 66027-5080
1	Director US Army TRADOC Systems Analysis Activity ATTN: ATAA-SL White Sands Missile Range, NM 88002	1	Commander Radford Army Ammo Plant ATTN: SMCRA-QA/HI LIB Radford, VA 24141
1	Commander US Army Training & Doctrine Command ATTN: ATCD-MA/ MAJ Williams Fort Monroe, VA 23651	1	Commander US Army Foreign Science & Technology Center ATTN: AMXST-MC-3 220 Seventh Street, NE Charlottesville, VA 22901
1	Commander US Army Research Office ATTN: Tech Library P. O. Box 12211 Research Triangle Park, NC 27709-2211	1	Commandant US Army Field Artillery Center & School ATTN: ATSF-CO-MW, B. Willis Ft. Sill, OK 73503
1	Commander US Army Belvoir Research & Development Ctr ATTN: STRBE-WC Tech Library (Vault) Bldg 315 Fort Belvoir, VA 22060-5606	1	Commander US Army Development and Employment Agency ATTN: MODE-TED-SAB Fort Lewis, WA 98433
		1	Office of Naval Research ATTN: Code 473, R. S. Miller 800 N. Quincy Street Arlington, VA 22217
		2	Commander Naval Sea Systems Command ATTN: SEA 62R SEA 64 Washington, DC 20362-5101

DISTRIBUTION LIST

<u>No. Of Copies</u>	<u>Organization</u>	<u>No. Of Copies</u>	<u>Organization</u>
1	Commander Naval Air Systems Command ATTN: AIR-954-Tech Lib Washington, DC 20360	3	Commander Naval Weapons Center ATTN: Code 388, R. L. Derr C. F. Price T. Boggs
1	Assistant Secretary of the Navy (R, E, and S) ATTN: R. Reichenbach Room 5E787 Pentagon Bldg. Washington, DC 20350		Info. Sci. Div. China Lake, CA 93555-6001
1	Naval Research Lab Tech Library Washington, DC 20375	1	Superintendent Naval Postgraduate School Dept. of Mechanical Engineering Code 1424 Library Monterey, CA 93943
2	Commander US Naval Surface Weapons Center ATTN: J. P. Consaga C. Gotzmer Silver Spring, MD 20902-5000	1	Program Manager AFOSR Directorate of Aerospace Sciences ATTN: L. H. Caveny Bolling AFB, DC 20332
3	Naval Surface Weapons Center ATTN: S. Jacobs/R10 Code 730 K. Kim/Code R-13 R. Bernecker/Code R-13 Silver Spring, MD 20902-5000	3	Commander Naval Ordnance Station ATTN: J. Birkett D. Brooks Tech Library Indian Head, MD 20640
5	Commander Naval Surface Weapons Center ATTN: Code G33, J. L. East W. Burrell J. Johndrow Code G23, D. McClure Code DX-21 Tech Lib Dahlgren, VA 22448-5000	1	HQ AFSC/SDOA Andrews AFB, MD 20334
2	Commander Naval Underwater Systems Center Energy Conversion Dept. ATTN: CODE 5B331, R. S. Lazar Tech Lib Newport, RI 02840	1	AFRPL/DY, Stop 24 ATTN: J. N. Levine/DYCR Edwards AFB, CA 93523-5000
		1	AFRPL/TSTL (Tech Library) Stop 24 Edwards AFB, CA 93523-5000
		1	AFRPL/MKPB, Stop 24 ATTN: B. Goshgarian Edwards AFB, CA 93523-5000
		1	AFFTC ATTN: SSD-Tech Lib Edwards AFB, CA 93523

DISTRIBUTION LIST

<u>No. Of Copies</u>	<u>Organization</u>	<u>No. Of Copies</u>	<u>Organization</u>
1	AFATL/DLYV ATTN: George C. Crews Eglin AFB, FL 32542-5000	1	Calspan Corporation ATTN: Tech Library P. O. Box 400 Buffalo, NY 14225
1	AFATL/DLJE Eglin AFB, FL 32542-5000	1	General Electric Company Armament Systems Dept. ATTN: M. J. Bulman, Room 1311 Lakeside Avenue Burlington, VT 05401
1	Air Force Armament Lab. AFATL/DLODL Eglin AFB, FL 32542-5000	1	IITRI ATTN: M. J. Klein 10 W. 35th Street Chicago, IL 60616
1	AFWL/SUL Kirtland AFB, NM 87117	1	Hercules Powder Co. Allegany Ballistics Laboratory ATTN: R. B. Miller P. O. Box 210 Cumberland, MD 21501
10	Central Intelligence Agency Office of Central Reference Dissemination Branch Room GE-47 HQS Washington, DC 20502	1	Hercules, Inc Bacchus Works ATTN: K. P. McCarty P. O. Box 98 Magna, UT 84044
1	Central Intelligence Agency ATTN: Joseph E. Backofen HQ Room 5F22 Washington, DC 20505	2	Director Lawrence Livermore National Laboratory ATTN: M. S. L-355, A. Buckingham P. O. Box 808 Livermore, CA 94550
1	General Applied Sciences Lab ATTN: J. Erdos Merrick & Stewart Avenues Westbury, NY 11590	1	Lawrence Livermore National Laboratory ATTN: M. S. L-355 M. Finger P. O. Box 808 Livermore, CA 94550
1	Aerodyne Research, Inc. Bedford Research Park ATTN: V. Yousefian Bedford, MA 01730	1	Olin Corporation Badger Army Ammunition Plant ATTN: R. J. Thiede Baraboo, WI 53913
1	Aerojet Solid Propulsion Co. ATTN: P. Micheli Sacramento, CA 95813		
1	Atlantic Research Corporation ATTN: M. K. King 5390 Cheorokee Avenue Alexandria, VA 22314		
1	AVCO Everett Rsch Lab ATTN: D. Stickler 2385 Revere Beach Parkway Everett, MA 02149		

DISTRIBUTION LIST

<u>No. Of Copies</u>	<u>Organization</u>	<u>No. Of Copies</u>	<u>Organization</u>
1	Olin Corp/Smokeless Powder Operations R&D Library ATTN: V. McDonald P.O. Box 222 St. Marks, FL 32355	1	Thiokol Corporation Wasatch Division ATTN: Tech Library P. O. Box 524 Brigham City, UT 84302
1	Paul Gough Associates, Inc. ATTN: P. S. Gough P. O. Box 1614, 1048 South St. Portsmouth, NH 03801	2	Thiokol Corporation Elkton Division ATTN: R. Biddle Tech Lib. P. O. Box 241 Elkton, MD 21921
1	Physics International Company ATTN: Library H. Wayne Wampler 2700 Merced Street San Leandro, CA 94577	1	Universal Propulsion Company ATTN: H. J. McSpadden Black Canyon Stage 1 Box 1140 Phoenix, AZ 85029
2	Rockwell International Rocketdyne Division ATTN: BA08 J. E. Flanagan J. Gray 6633 Canoga Avenue Canoga Park, CA 91304	2	United Technologies Chemical Systems Division ATTN: R. Brown Tech Library P. O. Box 358 Sunnyvale, CA 94086
1	Princeton Combustion Research Lab., Inc. ATTN: M. Summerfield 475 US Highway One Monmouth Junction, NJ 08852	1	Veritay Technology, Inc. 4845 Millersport Hwy. P. O. Box 305 East Amherst, NY 14051-0305
1	Science Applications, Inc. ATTN: R. B. Edelman 23146 Cumorah Crest Woodland Hills, CA 91364	1	Battelle Memorial Institute ATTN: Tech Library 505 King Avenue Columbus, OH 43201
3	Thiokol Corporation Huntsville Division ATTN: D. Flanigan R. Glick Tech Library Huntsville, AL 35807	1	Brigham Young University Dept. of Chemical Engineering ATTN: M. Beckstead Provo, UT 84601
1	Scientific Research Assoc., Inc. ATTN: H. McDonald P.O. Box 498 Glastonbury, CT 06033	1	California Institute of Tech 204 Karman Lab Main Stop 301-46 ATTN: F. E. C. Culick 1201 E. California Street Pasadena, CA 91109

DISTRIBUTION LIST

<u>No. Of Copies</u>	<u>Organization</u>	<u>No. Of Copies</u>	<u>Organization</u>
1	California Institute of Tech Jet Propulsion Laboratory ATTN: L. D. Strand 4800 Oak Grove Drive Pasadena, CA 91103	1	Massachusetts Institute of Technology Dept of Mechanical Engineering ATTN: T. Toong 77 Massachetts Avenue Cambridge, MA 02139
1	Professor Herman Krier Dept of Mech/Indust Engr University of Illinois 144 MEB; 1206 N. Green St. Urbana, IL 61801	1	Pennsylvania State University Dept. Of Mechanical Engineering ATTN: K. Kuo University Park, PA 16802
1	University of Minnesota Dept. of Mechanical Engineering ATTN: E. Fletcher Minneapolis, MN 55455	1	University of Michigan Gas Dynamics Lab Aerospace Engr Bldg ATTN: Dr. G. M. Faeth Ann Harbor, MI 48109-2140
1	University of Massachusetts Dept. of Mechanical Engineering ATTN: K. Jakus Amherst, MA 01002	1	Purdue University School of Mechanical Engineering ATTN: J. R. Osborn TSPC Chaffee Hall West Lafayette, IN 47906
1	Case Western Reserve University Division of Aerospace Sciences ATTN: J. Tien Cleveland, OH 44135	1	SRI International Propulsion Sciences Division ATTN: Tech Library 333 Ravenswood Avenue Menlo Park, CA 94025
3	Georgia Institute of Tech School of Aerospace Eng. ATTN: B. T. Zinn E. Price W. C. Strahle Atlanta, GA 30332	1	Director Los Alamos National Lab ATTN: M. Division, B. Craig T-3 MS B216 Los Alamos, NM 87545
1	Johns Hopkins University Applied Physics Laboratory Chemical Propulsion Information Agency ATTN: T. Christian Johns Hopkins Road Laurel, MD 20707	1	Stevens Institute of Technology Davidson Laboratory ATTN: R. McAlevy, III Castle Point Station Hoboken, NJ 07030

DISTRIBUTION LIST

<u>No. Of Copies</u>	<u>Organization</u>	<u>No. Of Copies</u>	<u>Organization</u>
1	Rutgers University Dept. of Mechanical and Aerospace Engineering ATTN: S. Temkin University Heights Campus New Brunswick, NJ 08903		
1	University of Southern California Mechanical Engineering Dept. ATTN: OHE200, M. Gerstein Los Angeles, CA 90007		
2	University of Utah Dept. of Chemical Engineering ATTN: A. Baer G. Flandro Salt Lake City, UT 84112		
1	Washington State University Dept. of Mechanical Engineering ATTN: C. T. Crowe Pullman, WA 99163		

Aberdeen Proving Ground

Dir, USAMSAA
ATTN: AMXSY-D
AMXSY-MP, H. Cohen

Cdr, USATECOM
ATTN: AMSTE-TO-F

Cdr, USACSTA
ATTN: S. Walton
G. Rice
D. Lacey
C. Herud

Dir, HEL
ATTN: J. Weisz

Cdr, CRDC, AMCCOM
ATTN: SMCCR-RSP-A
SMCCR-MU
SMCCR-SPS-IL

USER EVALUATION SHEET/CHANGE OF ADDRESS

This Laboratory undertakes a continuing effort to improve the quality of the reports it publishes. Your comments/answers to the items/questions below will aid us in our efforts.

1. BRL Report Number _____ Date of Report _____

2. Date Report Received _____

3. Does this report satisfy a need? (Comment on purpose, related project, or other area of interest for which the report will be used.) _____

4. How specifically, is the report being used? (Information source, design data, procedure, source of ideas, etc.) _____

5. Has the information in this report led to any quantitative savings as far as man-hours or dollars saved, operating costs avoided or efficiencies achieved, etc? If so, please elaborate. _____

6. General Comments. What do you think should be changed to improve future reports? (Indicate changes to organization, technical content, format, etc.) _____

CURRENT
ADDRESS

Name _____
Organization _____
Address _____
City, State, Zip _____

7. If indicating a Change of Address or Address Correction, please provide the New or Correct Address in Block 6 above and the Old or Incorrect address below.

OLD
ADDRESS

Name _____
Organization _____
Address _____
City, State, Zip _____

(Remove this sheet, fold as indicated, staple or tape closed, and mail.)

----- FOLD HERE -----

Director
US Army Ballistic Research Laboratory
ATTN: DRXBR-OD-ST
Aberdeen Proving Ground, MD 21005-5066




NO POSTAGE
NECESSARY
IF MAILED
IN THE
UNITED STATES

OFFICIAL BUSINESS
PENALTY FOR PRIVATE USE, \$300

BUSINESS REPLY MAIL
FIRST CLASS PERMIT NO 12062 WASHINGTON, DC
POSTAGE WILL BE PAID BY DEPARTMENT OF THE ARMY

Director
US Army Ballistic Research Laboratory
ATTN: DRXBR-OD-ST
Aberdeen Proving Ground, MD 21005-9989



----- FOLD HERE -----

